



# bioavid hook line evaluation card

bioavid lateral flow test kits with hook line – BLH-series  
Incubation time: 5 min



	T-line test line	H-line hook line	C-line control line	
Negative	-	+	+	
Positive	+	+	+	
Positive	+	+/-	+	
High Positive higher dilution to be tested	-	-	+	
Invalid	+/-	+/-	-	

08/2022

